

 **BD Rhapsody™ System**  
Single-Cell Capture and cDNA Synthesis  
with  
BD Rhapsody™ Single-Cell Analysis System  
Protocol

## Copyrights

No part of this publication may be reproduced, transmitted, transcribed, stored in retrieval systems, or translated into any language or computer language, in any form or by any means: electronic, mechanical, magnetic, optical, chemical, manual, or otherwise, without prior written permission from BD.

The information in this guide is subject to change without notice. BD reserves the right to change its products and services at any time. Although this guide has been prepared with every precaution to ensure accuracy, BD assumes no liability for any errors or omissions, nor for any damages resulting from the application or use of this information. BD welcomes customer input on corrections and suggestions for improvement.

## Trademarks

BD, the BD Logo, BD Rhapsody and Pharmingen are trademarks of Becton, Dickinson and Company or its affiliates. All other trademarks are the property of their respective owners. © 2022 BD. All rights reserved.

For US patents that may apply, see [bd.com/patents](https://www.bd.com/patents).

## Regulatory information

For Research Use Only. Not for use in diagnostic or therapeutic procedures.

## History

Revision	Date	Change made
Doc ID: 210966 Rev. 1.0	2018-07	Initial release.
23-22951-00 (Doc ID: 210966 Rev. 2.0)	2020-07	Assigned new 23 document part number, updated catalog number for reverse transcriptase, recommended thermomixer instead of heat block for Exonuclease I inactivation.
23-22951(01) (Doc ID: 210966 Rev. 3.0)	2022-01	Added BD Rhapsody™ Enhanced Cell Capture Beads and part numbers.
23-22951(02)	2022-11	Updated for BD Rhapsody™ Enhanced Cell Capture Beads v2.0. Removed part numbers.  Added note at the end of the <a href="#">Washing BD Rhapsody™ Enhanced Cell Capture Beads</a> workflow: <b>Note:</b> If performing the TCR and/or BCR assay, stop here and proceed with respective protocol to continue cDNA synthesis.

# Contents

---

<b>Introduction</b> .....	<b>4</b>
Required and recommended materials .....	4
Required reagents .....	4
Recommended consumables .....	5
Required equipment .....	6
Best practices .....	7
Before you begin .....	7
<b>Priming and treating the BD Rhapsody™ Cartridge</b> .....	<b>8</b>
<b>Staining cells with viability markers</b> .....	<b>8</b>
<b>Counting and preparing a single-cell suspension for cartridge loading</b> .....	<b>9</b>
<b>Loading cells in the cartridge</b> .....	<b>10</b>
<b>Preparing BD Rhapsody™ Enhanced Cell Capture Beads</b> .....	<b>10</b>
<b>Loading and washing BD Rhapsody™ Enhanced Cell Capture Beads</b> .....	<b>11</b>
<b>Lysing cells</b> .....	<b>12</b>
<b>Retrieving BD Rhapsody™ Enhanced Cell Capture Beads</b> .....	<b>12</b>
<b>Washing BD Rhapsody™ Enhanced Cell Capture Beads</b> .....	<b>13</b>
<b>Performing reverse transcription</b> .....	<b>14</b>
<b>Treating the sample with Exonuclease I</b> .....	<b>15</b>
<b>Troubleshooting</b> .....	<b>16</b>

## Introduction

---

This protocol describes cell loading in the BD Rhapsody™ Cartridge and single-cell capture with the BD Rhapsody™ Single-Cell Analysis System with the BD Rhapsody™ Scanner.

For complete instrument procedures and safety information, see the *BD Rhapsody™ Single-Cell Analysis System Instrument User Guide*.

### Required and recommended materials

For a complete list of materials, see the instrument user guide.

### Required reagents

Material	Supplier	Catalog no.
BD Rhapsody™ Enhanced Cartridge Reagent Kit	BD Biosciences	664887
BD Rhapsody™ Cartridge Kit	BD Biosciences	633733
BD Rhapsody™ cDNA Kit	BD Biosciences	633773
Absolute ethyl alcohol, molecule biology grade	Major supplier	–
Nuclease-free water	Major supplier	–
Calcein AM <sup>a</sup>	Thermo Fisher Scientific	C1430
DRAQ7™ <sup>a</sup>	BD Pharmingen™	564904
Dimethyl sulfoxide (DMSO)	Major supplier	–
70% ethyl alcohol or 70% isopropyl alcohol <sup>b</sup>	Major supplier	–

a. Protect Calcein AM and DRAQ7 from light. Avoid multiple freeze-thaw cycles of Calcein AM. See manufacturer's storage recommendations.

b. To clean the BD Rhapsody™ Express instrument and the BD Rhapsody™ Scanner, see the *BD Rhapsody™ Single-Cell Analysis System Installation and Maintenance Guide*. Instead of 70% alcohol, 10% (v/v) bleach can be used.

**Recommended consumables**

<b>Material</b>	<b>Supplier</b>	<b>Catalog no.</b>
Gilson™ PIPETMAN™ Tipack™ Filter Tips, 100-1200 µL for BD Rhapsody™ P1200M pipette	Thermo Fisher Scientific	F171803G
Gilson™ PIPETMAN™ Tipack™ Filter Tips, 500-5000 µL for BD Rhapsody™ P5000M pipette	Thermo Fisher Scientific	F161370G
Falcon® Tube with Cell Strainer Cap	Corning	352235
DNA LoBind® Tubes, 1.5-mL	Eppendorf	30108051
DNA LoBind® Tubes, 5.0-mL <sup>a</sup>	Eppendorf	30108310
Low-retention, filtered pipette tips (10 µL, 200 µL, 1000 µL)	Major supplier	–
INCYTO™ disposable hemocytometer	INCYTO	DHC-N01-5
Premoistened cleaning wipes with 70% ethyl alcohol or 70% isopropyl alcohol.	Major supplier	–
Lint-free wipes	Major supplier	–
a. These are the Bead Retrieval Tubes to be used with the BD Rhapsody™ Express instrument.		

**Required equipment**

Material	Supplier	Catalog no.
BD Rhapsody™ Scanner <sup>a</sup>	BD Biosciences	633701
BD Rhapsody™ Express Instrument <sup>a</sup>	BD Biosciences	633702
Hemocytometer Adapter <sup>a</sup>	BD Biosciences	633703
BD Rhapsody™ P1200M pipette <sup>a</sup>	BD Biosciences	633704
BD Rhapsody™ P5000M pipette <sup>a</sup>	BD Biosciences	633705
Large magnetic separation stand	V&P Scientific, Inc.	VP 772FB-1
Clear acrylic cylinder adapter for 15-mL tube adapter <sup>b</sup>	V&P Scientific, Inc.	VP 772FB-1A
Microcentrifuge for 1.5–2.0-mL tubes	Major supplier	–
Centrifuge and rotor with adapters for 5-mL Falcon <sup>®</sup> tubes and 15-mL tubes.	Major supplier	–
Eppendorf ThermoMixer <sup>®</sup> C	Eppendorf	5382000023
SmartBlock™ Thermoblock 1.5-mL	Eppendorf	5360000038
Water bath OR incubator at 37 °C	Major supplier	–
Pipettes (P10, P20, P200, P1000)	Major supplier	–
Vortexer	Major supplier	–
Digital timer	Major supplier	–
6-Tube Magnetic Separation Rack for 1.5-mL tubes Or, 12-Tube Magnetic Separation Rack Or, Invitrogen™ DynaMag™-2 Magnet	New England Biolabs  New England Biolabs  Thermo Fisher Scientific	S1506S  S1509S  12321D
a. Part of the BD Rhapsody™ Single-Cell Analysis system. b. Holds 5-mL LoBind tube in magnet.		

## Best practices

- Always use low-retention filtered pipette tips and LoBind tubes.
- Perform single-cell capture and cDNA synthesis in a pre-amplification workspace.
- Prepare cells as close to cell loading as possible. Keep the other reagents, including Sample Buffer, on ice unless instructed otherwise.
- Change pipette tips before every pipetting step.
- To ensure an air-tight seal with the BD Rhapsody™ P1200M and P5000M pipettes, hold the pipette with one hand and slightly twist the pipette to firmly seat a pipette tip on the pipette shaft.

## Before you begin

- Thaw Calcein AM. Once at room temperature (15–25 °C), resuspend Calcein AM in 503.0 µL DMSO for a final stock concentration of 2 mM. Follow the manufacturer's instructions and protect from light.
- Thaw reagents (not enzymes) in the BD Rhapsody™ cDNA Kit at room temperature (15–25 °C), and then place on ice. Keep enzymes at –25 °C to –15 °C.
- Place on ice the following components of the BD Rhapsody™ Enhanced Cartridge Reagent Kit:
  - Sample Buffer
  - 1 M DTT
  - Lysis Buffer
  - BD Rhapsody™ Enhanced Cell Capture Beads
- Ensure that the SmartBlock™ Thermoblock 1.5 mL or equivalent is installed on the thermomixer and is set to 37 °C for 20 minutes.
- Set an additional thermomixer to 80 °C if available.
- Prepare a single-cell suspension. See *Preparing Single-Cell Suspensions Protocol*.
- If your biological sample contains red blood cell contamination, red blood cell lysis is required. See *Preparing Single-Cell Suspensions Protocol*.
- Visually inspect the Lysis Buffer for any precipitation. If precipitation is present, incubate the Lysis Buffer at room temperature for 1 hour. Invert to mix, but do not vortex. Once the solution is clear, place the Lysis Buffer on ice.
- Open the tube while holding the DTT tube vertically. The solution is overlain with an inert/non-oxygen-containing gas and a non-vertical tube will allow the inert gas to pour off. If not loading four cell cartridges at the same time, after opening the DTT tube once, seal and store the tube at –20 °C.

## Priming and treating the BD Rhapsody™ Cartridge

Prime and treat the BD Rhapsody™ Cartridge. For detailed instructions, see the instrument user guide.

Express instrument slider	Position
Front	Waste
Side	0

Step no.	Material to load	Volume (μL)	P1200M pipette mode	Incubation at room temperature
1	100% ethyl alcohol	700	Prime/Treat	—
2	Air	700	Prime/Treat	—
3	Room temp. Cartridge Wash Buffer 1	700	Prime/Treat	1 min
4	Air	700	Prime/Treat	—
5	Room temp. Cartridge Wash Buffer 1	700	Prime/Treat	10 min
6	Air	700	Prime/Treat	—
7	Room temp. Cartridge Wash Buffer 2	700	Prime/Treat	≤4 hr

## Staining cells with viability markers

Protect Calcein AM and DRAQ7™ from light until ready to use.

- 1 If cells are not resuspended in cold Sample Buffer, centrifuge the cell suspension at  $400 \times g$  for 5 minutes, aspirate the supernatant, and leave ~20 μL of residual supernatant. Add up to 620 μL total volume of cold Sample Buffer.

Performance might be impacted if samples are not in Sample Buffer. For rare samples that are not resuspended in Sample Buffer before cell loading, proceed at your own risk, or contact tech support.

- 2 Add 3.1 μL of 2 mM Calcein AM and 3.1 μL of 0.3 mM DRAQ7™ to 620 μL of cell suspension (1:200 dilution) in cold Sample Buffer.
- 3 Gently pipet-mix.
- 4 Incubate at 37 °C in the dark for 5 minutes.
- 5 Filter cells through Falcon® Tube with Cell Strainer Cap.

For low-abundance or low-volume samples, filtering is optional at this step. We recommend filtering the final sample before loading cells into the cartridge.

- 6 Count cells immediately using the scanner.
  - a Ensure cells are well suspended by gently pipet-mixing.
  - b Pipet 10 μL into the INCYTO™ disposable hemocytometer.

Keep the remaining cells on ice, and protect from light.



## Counting and preparing a single-cell suspension for cartridge loading

---

For detailed instructions on counting cells with the BD Rhapsody™ Scanner, see the instrument user guide.

- 1 Insert the hemocytometer into the Hemocytometer Adapter, and tap **Scan**.
- 2 Place the adapter on the scanner tray, and tap **Continue**.
- 3 Select **Hemocytometer** for the protocol, and select or enter the experiment name, sample name, and user.
- 4 Tap **Side A** or **Side B**, then **Start Side A Scan** or **Start Side B Scan (Cell Count)**.
- 5 After the scan is complete, tap **OK**.
- 6 Tap **Scan**, and enter a new sample name to scan the other side of the hemocytometer. Repeat **steps 4–5**, or tap **Eject** and remove the adapter. Tap **Done**.
- 7 Tap **Analysis** and experiment name to view the **total cell concentration** and **cell viability**.
- 8 Proceed as follows:
  - If the cell concentration is  $\leq 1,000$  cells/ $\mu\text{L}$ , proceed to **step 9**.
  - If the cell concentration is  $> 1,000$  cells/ $\mu\text{L}$ , dilute the cell suspension in cold Sample Buffer to  $\sim 200$ – $800$  cells/ $\mu\text{L}$ . Repeat **steps 1–7**, and then **step 9**.
- 9 Tap **Prepare** to display the Samples Calculator screen.
- 10 Dispose of the hemocytometer.

Minimize the time between cell pooling and single-cell capture.
- 11 Use the Samples Calculator to obtain stock cell and buffer volumes from the scanner to prepare a cell suspension of  $650 \mu\text{L}$ . See the instrument user guide.
- 12 Prepare the cell suspension according to the displayed volumes on the scanner.

Ensure the stock solution of each sample is well suspended by gently pipet-mixing before pooling.
- 13 If the samples were not filtered before counting cells, filter through a Falcon® Tube with Cell Strainer Cap.

## Loading cells in the cartridge

1 Load the cartridge with materials listed using the P1200M pipette:

Material to load	Volume (μL)	Pipette mode
Air	700	Prime/Treat
<ul style="list-style-type: none"> <li>Set P1200M pipette to <b>Cell Load</b> mode.</li> <li>Pipet-mix the cell suspension with a manual P1000 pipette.</li> </ul>		
Cell suspension	575	<b>Cell Load<sup>a</sup></b>
<p>a. Press button once to aspirate 40 μL air, and then immerse tip in cell suspension. Press button again to aspirate 575 μL of cold cell suspension. Dispense 615 μL of air and cell suspension.</p>		

Air bubbles that might appear at the inlet or outlet of the cartridge do not affect cartridge performance.

- If necessary, wipe condensation from the top cartridge surface for optimal scanning.
- Incubate at room temperature (15–25 °C) for 15 minutes. To incubate the cartridge on the scanner, enter a time delay of 15 minutes before tapping **Start Cell Load Scan** (Cell Load step).
- During the 15-minute incubation, prepare the BD Rhapsody™ Enhanced Cell Capture Beads. See [Preparing BD Rhapsody™ Enhanced Cell Capture Beads](#) in the following section.
- Image the cells in the cartridge. Perform the scanner step: **Cell Load**. For more information, see the instrument user guide.
- After the scan is complete, tap **OK** and **Eject**. Remove the cartridge and tap **Scan** or **Done**. After the scan, confirm the analysis is running.

## Preparing BD Rhapsody™ Enhanced Cell Capture Beads

Keep BD Rhapsody™ Enhanced Cell Capture Beads on ice before use.

For maximum recovery, do not vortex samples containing BD Rhapsody™ Enhanced Cell Capture Beads. Gently mix suspensions with BD Rhapsody™ Enhanced Cell Capture Beads by pipette only.

Use low-retention pipette tips and LoBind tubes. Keep the beads cold, and pipet-mix only.

- Place the BD Rhapsody™ Enhanced Cell Capture Beads tube on the magnet for 1 minute, and remove the storage buffer.
- Remove the tube from the magnet, and pipet 750 μL of cold Sample Buffer into the tube.
- Pipet-mix, and place on ice.
- After the Cell Load scan and analysis is running, proceed to [Loading and washing BD Rhapsody™ Enhanced Cell Capture Beads](#) in the following section.

## Loading and washing BD Rhapsody™ Enhanced Cell Capture Beads

- 1 Place the cartridge on the Express instrument.
- 2 Set the P1200M pipette to **Prime/Treat** mode.
- 3 Load the cartridge with the materials listed using the P1200M pipette:

Material to load	Volume (µL)	Pipette mode
Air	700	Prime/Treat
<ul style="list-style-type: none"> <li>• Set P1200M pipette to <b>Bead Load</b> mode.</li> <li>• Use a manual P1000 to gently pipet-mix the beads in cold Sample Buffer. Immediately load.</li> </ul>		
BD Rhapsody™ Enhanced Cell Capture Beads	630	<b>Bead Load</b>

- 4 Incubate the cartridge at room temperature (15–25 °C) for 3 minutes.
- 5 Perform the scanner step: **Bead Load**.
- 6 Place the cartridge on the Express instrument.
- 7 Set the P1200M pipette to **Wash** mode.
- 8 Load the cartridge with the materials listed using the P1200M pipette:

Material to load	Volume (µL)	Pipette mode <sup>a</sup>
Air	700	Wash
Cold Sample Buffer	700	Wash
Air	700	Wash
Cold Sample Buffer	700	Wash

a. Press button once to aspirate 720 µL air or reagent. Insert the tip into the cartridge, and press button once to dispense 700 µL air or liquid. Remove pipette tip, and press button once to dispense remaining 20 µL of air or liquid.

- 9 Perform the scanner step: **Bead Wash**.

## Lysing cells

---

Avoid bubbles.

- 1 Add 75.0 µL of 1 M DTT to one 15-mL Lysis Buffer bottle. Check Add DTT box.

Use the Lysis Buffer with DTT ≤24 hours, and then discard.

- 2 Briefly vortex the lysis mix, and place on ice.
- 3 Place the cartridge on the Express instrument.
- 4 Move the left slider to LYSIS.
- 5 Set the P1200M pipette to **Lysis** mode.
- 6 Load the cartridge with the materials listed using the P1200M pipette:

Material to load	Volume (µL)	Pipette mode
Lysis Buffer with DTT	550	Lysis

- 7 Incubate at room temperature (15–25 °C) for 2 minutes.

Maintain the recommended lysis time for best performance.

## Retrieving BD Rhapsody™ Enhanced Cell Capture Beads

---

- 1 Place the 5-mL LoBind tube in the Express instrument drawer.
- 2 Ensure that the P5000M pipette is set to **Retrieval** mode.
- 3 Move the front slider to BEADS on the Express instrument.
- 4 Move the left slider to RETRIEVAL.
- 5 Leave the Retrieval magnet in the down position for 30 seconds.
- 6 Aspirate 5,000 µL of Lysis Buffer with DTT using the P5000M pipette.
- 7 Press down on the P5000M pipette to seal against the gasket.
- 8 Move the left slider to the middle position (0), and immediately load 4,950 µL of Lysis Buffer with DTT.
- 9 Remove the pipette from the gasket, and purge the tip.
- 10 Move the front slider to OPEN, and place the 5-mL LoBind tube on the large magnet with the 15-mL tube adapter for 1 minute.
- 11 During the 1-minute incubation, perform the scanner step: Retrieval.
- 12 Immediately proceed to [Washing BD Rhapsody™ Enhanced Cell Capture Beads](#) in the following section.
- 13 Appropriately dispose of the cartridge, Waste Collection Container, and Lysis Buffer with DTT.



**Biological hazard.** All surfaces that come in contact with biological specimens can transmit potentially fatal disease. Use universal precautions when cleaning surfaces. Wear suitable protective clothing, eyewear, and gloves.

- 14 Clean the Express instrument with 10% bleach or 70% ethyl alcohol.

## Washing BD Rhapsody™ Enhanced Cell Capture Beads

---

- 1 After the 1-minute incubation, leaving the 5-mL tube containing retrieved BD Rhapsody™ Enhanced Cell Capture Beads on the large magnet, remove all but ~1 mL of supernatant without disturbing the beads.
- 2 Remove the tube from the magnet. Gently pipet-mix the beads and transfer them to a new 1.5-mL LoBind tube.
- 3 If there are still beads left in the 5-mL tube, add 0.5 mL of Lysis Buffer with DTT, rinse the 5-mL tube, and transfer to the 1.5-mL LoBind tube from the previous step.
- 4 Place the tube on the 1.5-mL magnet for  $\leq 2$  minutes. Remove the supernatant.  
  
Avoid leaving Lysis Buffer or bubbles in the tube. Lysis Buffer might cause the reverse transcription reaction to fail.
- 5 Remove the tube from the magnet, and pipet 1.0 mL of cold Bead Wash Buffer into the tube. Pipet-mix.
- 6 Place the tube on the 1.5-mL magnet for  $\leq 2$  minutes. Remove the supernatant.
- 7 Remove the tube from the magnet, and pipet 1.0 mL of cold Bead Wash Buffer into the tube. Pipet-mix, and place on ice.

Start reverse transcription  $\leq 30$  minutes after washing the retrieved BD Rhapsody™ Enhanced Cell Capture Beads with Bead Wash Buffer.

**Note:** If performing the TCR and/or BCR assay, stop here and proceed with respective protocol to continue cDNA synthesis.

## Performing reverse transcription

- 1 Ensure that the SmartBlock™ Thermoblock for ThermoMixer® C is at 37 °C.
- 2 In the pre-amplification workspace, pipet the following reagents into a new 1.5-mL LoBind tube on ice:

### cDNA mix

Component	For 1 library (µL)	For 1 library + 20% overage (µL)
RT Buffer	40.0	48.0
dNTP	20.0	24.0
RT 0.1 M DTT	10.0	12.0
Bead RT/PCR Enhancer	12.0	14.4
RNase Inhibitor	10.0	12.0
Reverse Transcriptase	10.0	12.0
Nuclease-Free Water	98.0	117.6
<b>Total</b>	<b>200.0</b>	<b>240.0</b>

- 3 Gently vortex mix, briefly centrifuge, and place back on ice.
- 4 Place the tube of washed BD Rhapsody™ Enhanced Cell Capture Beads on the 1.5-mL tube magnet for ≤2 minutes. Remove the supernatant.
- 5 Remove the tube from the magnet and pipet 200 µL of cDNA mix into the beads. Pipet-mix.  
Keep the prepared cDNA mix with beads on ice until the suspension is transferred in the next step.
- 6 Transfer the bead suspension to a new 1.5-mL LoBind tube.
- 7 Incubate the bead suspension on the SmartBlock™ Thermoblock for ThermoMixer® C at 1,200 rpm and 37 °C for 20 minutes.  
**Note:** Shaking is critical for this incubation.
- 8 During reverse transcription incubation, view the image analysis to see if the analysis metrics passed.
- 9 Place the tube on ice.

## Treating the sample with Exonuclease I

- 1 Set one thermomixer to 37 °C and a second thermomixer to 80 °C.

**Note:** Exonuclease I inactivation temperatures above 80 °C can result in the loss of AbSeq molecules, thus a heat block should not be used for this step. If only one thermomixer is available, allow it to equilibrate to 80 °C before starting the inactivation incubation.

- 2 In the pre-amplification workspace, pipet the following reagents into a new 1.5-mL LoBind tube on ice:

### Exonuclease I mix

Component	For 1 library (µL)	For 1 library + 20% overage (µL)
10X Exonuclease I Buffer	20.0	24.0
Exonuclease I	10.0	12.0
Nuclease-Free Water	170.0	204.0
<b>Total</b>	<b>200.0</b>	<b>240.0</b>

- 3 Gently vortex mix, briefly centrifuge, and place back on ice.
- 4 Place the tube of BD Rhapsody™ Enhanced Cell Capture Beads with cDNA mix on the 1.5-mL tube magnet for ≤2 minutes. Remove the supernatant.
- 5 Remove the tube from the magnet, and pipet 200 µL of Exonuclease I mix into the tube. Pipet-mix.
- 6 Incubate the bead suspension on the thermomixer at 1,200 rpm and 37 °C for 30 minutes.  
  
If only one thermomixer is available, allow it to equilibrate to 80 °C before starting the inactivation incubation. Place the samples on ice until that temperature is reached.
- 7 Incubate the bead suspension on the thermomixer (no shaking) at 80 °C for 20 minutes.  
  
Do not exceed this inactivation temperature and incubation time.
- 8 Place the tube on ice for ~1 minute.
- 9 Place the tube on the magnet for ≤1 minute until clear. Remove the supernatant.
- 10 Remove the tube from the magnet, and pipet 200 µL of cold Bead Resuspension Buffer into the tube. Pipet-mix.

**Stopping point:** Exonuclease I-treated beads can be stored at 2–8 °C for up to 3 months.

- 11 Proceed to library preparation. See the *Single-Cell Analysis Workflow with BD Rhapsody™ System*.

## Troubleshooting

---

For additional troubleshooting on scanning or cartridge loading, see the troubleshooting section in the instrument user guide.

For technical support, contact your local Field Application Specialist (FAS) or [scomix@bdscomix.bd.com](mailto:scomix@bdscomix.bd.com).

Observation	Possible causes	Recommended solutions
Reported viability from the BD Rhapsody™ Scanner suspected to be too high.	DRAQ7™ staining in the current protocol is optimized for cell lines. The optimal concentration of DRAQ7™ might be higher.	Before the BD Rhapsody™ experiment, optimize the DRAQ7™ concentration for your cell types according to the manufacturer's protocol. See the DRAQ7™ technical data sheet, <a href="http://bdbiosciences.com/ds/pm/tds/564904.pdf">bdbiosciences.com/ds/pm/tds/564904.pdf</a> .
No pellet after centrifuging cells or very few cells.	Rare or dilute sample.	After each centrifugation step, leave 50 µL of supernatant.



**Becton, Dickinson and Company**  
**BD Biosciences**  
2350 Qume Drive  
San Jose, California 95131 USA

[bdbiosciences.com](http://bdbiosciences.com)  
[scomix@bdscomix.bd.com](mailto:scomix@bdscomix.bd.com)