

Technical Data Sheet

APC/Myeloid-Cell Protein Panel

Product Information

Material Number:	572435
Size:	2 Tests
Vol. per Test:	NA
Reactivity:	Tested in Development: Human
Storage Buffer:	Lyophilized powder containing BSA and $\leq 0.25\%$ sodium azide.
Component:	51-9025300
Description:	APC/Myeloid-Cell Protein Panel
Size:	1 Test (2 ea)
Vol. per Test:	NA

Description

The BD® OMICS-One APC/Myeloid-Cell Protein Panel consists of 30 different specificities against major APC/Myeloid-Cell markers in a single tube. Designed and optimized to work on the BD Rhapsody™ System, the APC/Myeloid-Cell Protein Panel is tested to work seamlessly alongside the BD Rhapsody™ Whole Transcriptome Analysis (WTA) Assay, Targeted mRNA Assay, BD® Single-Cell Multiplexing Kit (SMK), BD® Intracellular CITE-seq (IC-AbSeq) Assay, and BD Rhapsody™ TCR/BCR Next Multiomic Assay for human. The individual antibodies were each conjugated to an oligonucleotide that contains a specific antibody barcode sequence flanked by a polyA tail on the 3' end and a common PCR handle (PCR primer binding site) on the 5' end. All AbSeq barcode sequences were generated in-silico with minimal sequence similarity to the human genomes, have low predicted secondary structure, and have high Hamming distance within the BD antibody-oligo portfolio, to allow for sequencing error correction and unique mapping. The polyA tail of the oligonucleotide allows the barcode sequence to be captured by BD Rhapsody™ Enhanced Cell Capture Beads. The 5' PCR handle allows for efficient library generation for various sequencing platforms. Each individual antibody exists at an optimal concentration within the 30-plex to enable superior target and population resolution.

The APC/Myeloid-Cell Protein Panel is designed with SMART technology. SMART technology helps lower sequencing cost while increasing data resolution by attenuating antibodies that target high-expressing primary markers and allowing reallocation of sequencing reads to markers expressed at lower levels. With SMART technology, markers low in expression can be quantified without having to do deeper sequencing and incurring high sequencing cost. The two specificities attenuated in the APC/Myeloid-Cell Protein Panel are CD45 and HLA-DR.

Preparation and Storage

Store at 2-8°C and protected from prolonged exposure to light.

Do not freeze.

Application Notes

Application

Single Cell 3' Sequencing	Qualified
---------------------------	-----------

Recommended Assay Procedure:

This reagent is provided lyophilized in a pre-titrated format.

1. Remove the BD® OMICS-One APC/Myeloid-Cell Protein Panel tube from the foil bag and bring up to room temperature for 5 minutes.
2. Make sure the pellet is located at the bottom of the tube. If not, briefly centrifuge to collect the contents at the tube bottom.
3. Add 35 μ L of nuclease-free water to the bottom of the tube and allow antibodies to reconstitute for 5 minutes at room temperature.
4. Place the reconstituted antibodies on ice until the cells are ready for staining.
Note: Reconstitute antibody immediately before cell staining. Prolonged incubation of reconstituted antibody may increase the non-specific background.
5. For BD® AbSeq Ab-Oligo drop-in of 60 plex or lower, prepare the BD® AbSeq labeling MasterMix in 1.5-mL LoBind tube on ice.
Note: For drop-in with more than 60 plex, reach out to technical support for calculation.

If sequential labeling with Sample Tags or no Sample Tags, prepare BD® AbSeq labeling MasterMix for drop-ins as follows:

Component	1 sample (μ L)	1 sample +	2 samples +
-----------	---------------------	------------	-------------

BD Biosciences

bdbiosciences.com
 United States Canada Europe Japan Asia Pacific Latin America/Caribbean
 877.232.8995 866.979.9408 32.2.400.98.95 0120.8555.90 65.6861.0633 55.11.5185.9995
 For country contact information, visit bdbiosciences.com/contact

Conditions: The information disclosed herein is not to be construed as a recommendation to use the above product in violation of any patents. BD Biosciences will not be held responsible for patent infringement or other violations that may occur with the use of our products. Purchase does not include or carry any right to resell or transfer this product either as a stand-alone product or as a component of another product. Any use of this product other than the permitted use without the express written authorization of Becton, Dickinson and Company is strictly prohibited.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Not for resale.
 © 2017 BD. BD, the BD Logo and all other trademarks are property of Becton, Dickinson and Company.



		30% overage (μL)	30% overage (μL)
Per BD® AbSeq Ab-Oligo	2.0	2.6	5.2
Total of BD® AbSeq Ab-Oligo	2.0 × N*	2.6 × N	5.2 × N
FBS† (catalog number 554656)	140 – (2.0 × N)	182 – (2.6 × N)	364 – (5.2 × N)
Total	140	182	364

If co-labeling with Sample Tags, prepare BD® AbSeq labeling MasterMix for drop-ins as follows:

Component	1 sample (μL)	1 sample + 30% overage (μL)	2 samples + 30% overage (μL)
Per BD® AbSeq Ab-Oligo	2.0	2.6	5.2
Total of BD® AbSeq Ab-Oligo	2.0 × N*	2.6 × N	5.2 × N
FBS† (catalog number 554656)	120 – (2.0 × N)	156 – (2.6 × N)	312 – (5.2 × N)
Total	120	156	312

* N = number of drop-in antibodies. N = 0 if there are no drop-in antibodies.

† FBS = BD Pharmingen™ Stain Buffer.

6. Pipet-mix the BD® AbSeq labeling MasterMix for drop-ins. Briefly centrifuge to collect the contents at the bottom, and place back on ice.
7. If sequential labeling with Sample Tags or no Sample Tags, for each sample, add 140 μL BD® AbSeq labeling MasterMix of drop-ins to the tube containing 35 μL reconstituted APC/Myeloid-Cell Protein Panel solution to make a total volume of 175 μL.
If co-labeling with Sample Tags, for each sample, add 120 μL BD® AbSeq labeling MasterMix of drop-ins and 20 μL Sample Tag to the tube containing 35 μL reconstituted APC/Myeloid-Cell Protein Panel solution to make a total volume of 175 μL.
8. Pipet-mix the mixture, briefly centrifuge to collect the contents at the tube bottom, and place back on ice.
9. Centrifuge cells at 400 × g for 5 minutes. If Fc Block is used, proceed to step 10. If Fc Block is not used, skip to step 11.
10. (Optional) For samples containing myeloid and B lymphocytes, BD Biosciences recommends blocking nonspecific Fc Receptor-mediated false-positive signals with Human BD Fc Block (catalog number 564220).
 - a. To perform blocking, pipet the Fc Block MasterMix into a new 1.5-mL LoBind tube on ice:

Component	1 sample (μL)*	1 sample + 20% overage (μL)
FBS† (catalog number 554656)	20.0	24.0
Fc Block‡ (catalog number 564220)	5.0	6.0
Total	25.0	30.0

* Sufficient for up to 1 million cells. To block more cells, adjust the volume.

† FBS = BD Pharmingen™ Stain Buffer.

‡ Fc Block = BD Pharmingen™ Human BD Fc Block.

- b. Pipet-mix the Fc Block MasterMix and briefly centrifuge. Place on ice.
 - c. Remove the supernatant from the cells without disturbing the pellet.
 - d. Resuspend the cells in 25 μL of Fc Block MasterMix.
 - e. Incubate the cells at room temperature (15°C to 25°C) for 10 minutes.
 - f. Add 175 μL of BD® AbSeq labeling MasterMix from Step 8 into the cell suspension. Pipet-mix and proceed to Step 12.
 11. Remove the supernatant from the cells without disturbing the pellet. Add 25 μL Stain Buffer (FBS) to the 175 μL of BD® AbSeq labeling MasterMix from Step 8 to make a total volume of 200 μL. Resuspend the cell pellet in 200 μL total volume. Pipet-mix.
 12. Transfer the cells with BD® AbSeq labeling MasterMix into a new 5-mL polystyrene Falcon tube.
 13. Stain the cells on ice for 30 minutes.
 14. Add 3–4 mL Stain Buffer (FBS) to labelled cells and pipet-mix.
 15. Centrifuge at 400 × g for 5 minutes.
 16. Uncap the tube and invert to decant supernatant into biohazardous waste. Keep the tube inverted and gently blot on a lint-free wiper to remove residual supernatant from tube rim.
 17. Repeat steps 14–16 twice more for a total of three washes.
 18. Resuspend the final washed cell pellet in 620 μL cold Sample Buffer from the BD Rhapsody™ Enhanced Cartridge Reagent V3 (catalog number 667052) and proceed to single cell capture with on-cartridge washing described in substeps a–c. Refer to the *BD Rhapsody™ HT Single-Cell Analysis System Single-Cell Capture and cDNA Synthesis Protocol* (Doc ID 23-24252) or *BD Rhapsody™ HT Xpress System Single-Cell Capture and cDNA Synthesis Protocol* (Doc ID 23-24253) for additional details.
- Note:** Perform on-cartridge washing after cell settling (8-minute incubation) as described in the following sub-steps.
- a. At the protocol section of “Loading cells in BD Rhapsody™ 8-Lane Cartridge”, after cell load, incubate the cartridge in the dark at

room temperature for 8 minutes.

- b. Place the cartridge on the BD Rhapsody™ HT Xpress and perform the On-Cartridge Wash steps as follows:

Material to load	Volume (µL)	1 lane	Pipette Mode
Air	380		Prime/Wash
Cold Sample Buffer	380		Prime/Wash
Air	380		Prime/Wash
Cold Sample Buffer	380		Prime/Wash

- c. (Optional) Perform the scanner step: Cell Load Scan, if using *BD Rhapsody™ HT Single-Cell Analysis System Single-Cell Capture and cDNA Synthesis Protocol* (Doc ID 23-24252). No need for 8-minute delay before scanning.

Warning: All biological specimens and materials are considered biohazardous. Handle as if capable of transmitting infection and dispose using proper precautions in accordance with federal, state, and local regulations. Never pipette by mouth. Wear suitable protective clothing, eyewear, and gloves.

List of all 30 Human AbSeq specificities included in the BD® OMICS-One T-Cell panel:

Specificity	Clone	Oligo ID	BD® AbSeq Barcode Sequence
CD103	BER-ACT	AHS0001	AAATAGTATCGAGCGTAGTTAAGTTGCGTAGCCGTT
CD274 (PD-L1)	MIH1	AHS0004	ATCGTAAGGCTCGTGGTTCGTAAGTAAGTTCGTATC
CD11b	M1/70	AHS0005	ATCGTTATTCTGTTGATGTTTCGCCGGTTTGAGTAGT
CD123	7G3	AHS0020	ACAGTTTAGTAGGACGTGAGGTATCGCGAGAATGCC
HLA-DR	G46-6	AHS0035	TGTTGGTTATTCTGTTAGTGCATCCGTTTGGGCGTGG
CD14	MPHI9	AHS0037	TGGCCCGTGGTAGCGCAATGTGAGATCGTAATAAGT
CD45	HI30	AHS0040	GTGCGAAATGGCGGAATGTTATCTGCGAATGTAGTC
CD33	WM53	AHS0044	GTGTTAGTGATTTGATAGGACGCGTTACGAGAGATT
CD80	L307.4	AHS0046	GAGGGTAACGGGTGTCCAAATATCGGCTGTGTAAGT
CD16	3G8	AHS0053	TAAATCTAATCGCGTAACATAACGGTGGGTAAGGT
CD64	10.1	AHS0055	TTGTGCGGCGTAGTATGGTTATCTCGAGTGAAAGTC
CD11c	B-LY6	AHS0056	ATGCGTTGCGAGAGATATGCGTAGGTTGCTGATTGG
CD163	GHI/61	AHS0062	TATTATGTGCGAACTATGGTATCCGTATTGAGGGCT
CD195 (CCR5)	2D7/CCR5	AHS0070	ATGGTTTAGTCGTACGTGGGTTTAGATTGGCGGTGC
CD206	19.2	AHS0072	GCTGGTTATCGTTTGAGAGTCGGTATGGAATGCGGT
CD32	FLI8.26	AHS0073	GGTTGTAGGTGCGGAATATAAGCGTCGTTGAGGTGT
CD273	MIH18	AHS0075	TGAGTAACCGTATGTAATCCGTAATCGTAGAAGCGC
CD141	1A4	AHS0083	TGGAAGTAAGTATGGGTCGGCGTAAATTGTGCGTGT
CD1	F10/21A3	AHS0088	ATAGATTACATTCGTTTAGCGTTGGGTTTCGGTCCGT
CD40	5C3	AHS0117	GGTGAATTGGGCTAGAACGTATATGCGGTAAGGCG
FCeR1a	AER-37	AHS0129	GATATGGCGTGATGGTAGGTTTCGGTTAAGTTAGCG
CD169	7-239	AHS0133	CATTAAAGCACGAAGGGTATAGGTAGGAACGGTTGGC
CD36	IVC7	AHS0135	AATTGTAGTAGTCCGGTGTATGTAGTAGTAGGCGTTT
CD115 (CSF1R)	9-4D2-1E4	AHS0136	CTGGTGGCGCGCAATTGTTTACGACATATAGGGTT
CD162	KPL-1	AHS0139	CCAGATAGGCGATAGTGTTTAGGAGCGATTAGTGTG
CD85K	ZM3.8	AHS0179	AGTAGTCGTAGTTGGCGTGAATTGGGCTTATATCTG
VISTA	MIH65.RMAB	AHS0187	ATCAGGGAATCTCGGTAAGTTAAACGTGTATAGTGC
CD15	W6D3	AHS0196	ATAGGCATGGACGACGTAGATAATAAGTGGCGGGTT
CD192 (CCR2)	LS132.1D9	AHS0208	CATGAGTGAGGCGATATAGTGAGCGGTTGTAGATT
CD116	Hgmcsfr1-M1	AHS0238	CTTAGTTGTAGGATCGAGAGTAGGTTGTGATTGCGT

Suggested Companion Products

Catalog Number	Name	Size	Clone
554656	Stain Buffer (FBS)	500 mL	(none)
564220	Human BD Fc Block™	0.25 mg	Fc1
633801	Whole Transcriptome Analysis (WTA) Amplification Kit	1 Each	(none)
633774	Targeted mRNA and AbSeq Amplification Kit	1 Each	(none)
667058	TCR/BCR Next Amplification Kit	1 Each	(none)
633773	cDNA Kit	1 Each	(none)
666262	8-Lane Cartridge	1 Each	(none)
667052	Enhanced Cartridge Reagent Kit V3	1 Each	(none)
633781	Hu Single Cell Sample Multiplexing Kit	1 Each	(none)
633849	Flex Single-Cell Multiplexing Kit A, Flex Sample Tag 1-6	1 Each	(none)

633850	Flex Single-Cell Multiplexing Kit B, Flex Sample Tag 7-12	1 Each	(none)
633851	Flex Single-Cell Multiplexing Kit C, Flex Sample Tag 13-18	1 Each	(none)
633852	Flex Single-Cell Multiplexing Kit D, Flex Sample Tag 19-24	1 Each	(none)
666625	BD® Rhapsody™ HT Xpress Package		(none)
633701	BD Rhapsody Scanner		(none)
570742	Intracellular AbSeq Buffer Kit	1 Each	(none)
570911	OMICS-Guard Sample Preservation Buffer	50 mL	(none)
570750	AbSeq Enhancer Kit	1 Each	(none)
570751	RNase Inhibitor	1 Each	(none)

Product Notices

1. Please refer to www.bdbiosciences.com/us/s/resources for technical protocols.
2. Please refer to <http://regdocs.bd.com> to access safety data sheets (SDS).
3. For U.S. patents that may apply, see bd.com/patents.
4. This reagent has been pre-diluted for use at the recommended Volume per Test. We typically use 1×10^6 cells in a 100- μ l experimental sample (a test).